

Technical Information

NTX-Urine

N-Terminal Telopeptide of Type I Collagen

IVD

Cat. No.: 9006
 Tests: 96
 Method: **ELISA**
 Range: 1 - 3000 nM BCE
 Sensitivity: 20 nM BCE (value not corrected for creatinine)
 Incubation time: 2 hours
 Sample volume: 25 µl
 Sample type: Urine, cell culture
 Sample preparation: Collect second void of morning urine or 24 hours urine without adding preservatives. Fasting sample collection is not required. Samples are stable at room temperature for 24 hours or refrigerated at 2-8 °C for up to 72 hours. Longer storage at -20°C. Stable up to 3 freeze and thaw cycles.

Reference values: nmol Bone Collagen Equivalents/l (nM BCE)
 per mmol Creatinine/l = nM BCE / mM Creatinine

	Mean	Range (± 2 SD)
Premenopausal women (25 - 49 yrs)	35	5 - 65
Men (31 - 87 yrs)	33	3 - 63

Reference values from children are available.

Species: Human, dog, porcine, rat, bovine and rhesus macaque

Intended use:

NTX-Urine test is used for the quantitative measured of cross-linked N-telopeptides of type I collagen in urine. The NTX molecule is bone specific due to the unique amino acid sequences of the cross-linked alpha-2 (I) N- telopeptide. Release of the NTX molecule is mediated by osteoclasts. Elevated levels of NTX indicate elevated bone resorption.

NTX-levels are used for:

- monitoring of antiresorptive therapy in osteoporosis
- Paget's disease.
- monitoring and early detection of bone metastasis.

For further information please contact / Für weitere Informationen wenden Sie sich bitte an /
Pour plus d'informations, veuillez contacter:

www.tecomedical.com



A EUROBIO SCIENTIFIC COMPANY

Switzerland / Headquarters

TECOmedical AG
Gewerbstrasse 10
4450 Sissach
Phone +41 61 985 81 00
Fax +41 61 985 81 09
Mail info@tecomedical.com

Germany

TECOmedical GmbH
Wasserbreite 57
32257 Bünde
Phone +49 52 23 985 99 99
Fax +49 52 23 985 99 98
Mail info@tecomedical.com

Austria

TECOmedical AG
Phone 0800 20 40 66
Fax 0800 20 40 55
Mail info@tecomedical.com